



GENERAL INFORMATION

- Category : Sterile sampling kits
- Family : Sani-S'wipe
- Lifespan: 12 months

TECHNICAL DESCRIPTION

This kit consists of a sterile bag with an integrated closure system and a biocide-free cloth for sampling various surfaces.



SPECIFIC INFORMATION

CONTAINER

Type of container :	Bag
Material :	Thermoplastic laminate
Color :	Transparent
Dimension :	190 x 292mm / 7.5 X 11.5 po
Thickness :	3.5 mil.in / 0.0889 mm / 88.9 micron
Total volume :	2100 ml / 70 oz
Functional volume :	1200 ml / 41 oz
Printing type :	Form
Opening system :	Laser scoring line
Closing system :	Attachment with 2 round wires
Sterile :	Yes
End of product life :	Garbage

INSERTION

Insertion type :	Blue Cloth
Description :	Non-woven textile composed of cellulose and polyester. Color: pale blue. Folded in 4. Dimensions 9" x 9" (22.9 cm x 22.9 cm). Size (total area): 81 sq. in. (522.5 cm ²). Free from biocides
Biodegradable :	Garbage

INSERTION

Insertion type :	
Description :	
Biodegradable :	

SOLUTION

Name :	SOL-NB
Type of solution :	Neutralizing
pH :	7.2±0.2

OTHER

AVAILABLE DOCUMENTS

Data Sheet	Certificate of Compliance
Certificate of Analysis	Safety Data Sheet (SDS)
Certificate of Sterility	Pyrogen Declaration
DNase/RNase	

PACKAGING INFORMATION

Outer box dimension : (W x D x H)	9.50 X 14.13 X 16.25
Box weight :	26.00 LB / 11.79 KG
Conditioning :	100 (1 x 100)
Storage condition :	Upon receipt, store between 2 °C and 8 °C

Reach out to us for additional resources, if applicable to this product.

DECLARATION

CFIA	LABPLAS sampling bags are a solution that may be used in the CFIA Preventive Control Plan (PCP) for the seven principles of the HACCP system. The PCP is a Canadian federal initiative, under the Safe Food for Canadians Regulations (SFCR).
EU	The materials used to manufacture LABPLAS sampling bags meet, where applicable, the Eu No10/2011 standards for food contact with respect to particle migration.
FDA	The plastic film used in the manufacture of LABPLAS sampling bags complies with 21 CFR 177.1520 [(c) 3.2c] of the Food and Drug Administration. Its use is limited to temperatures below 212 F according to Table 2 of 21 CFR 176.170 (c).
Biocide-free	Biocide-free material
Sterile	Sterility is provided by irradiation. The sterilization dose ensuring a sterility assurance level (SAL) of 10 ⁻³ has been established according to ISO 11137.

1951, rue Nobel, Sainte-Julie, Québec, Canada, J3E 1Z6

Tél : 1.450.649.7343 | Fax : 1.450.649.3113

labplas@labplas.com | www.labplas.com