

GENERAL INFORMATION

Category:
Sterile sampling bags

Family: Twirl'em

Lifespan: 5 years

TECHNICAL DESCRIPTION

Twirl'em bags have a practical and easy-to-use closing system. They are made of a flexible, strong and transparent plastic.



SPECIFIC INFORMATION

ITEM

| | |
|-----------------------|------------------------------------|
| ITEM | Bag |
| Material : | Polyethylene blend |
| Color : | Transparent |
| Dimension : | 140 x 305 mm / 5.5 X 12 po |
| Thickness : | 3 mil.in / 0.0762 mm / 76.2 micron |
| Total volume : | 1400 ml / 47 oz |
| Functional volume : | 860 ml / 29 oz |
| Printing type : | Writing area |
| Opening system : | Perforated line |
| Closing system : | Attachment with 2 round wires |
| Sterile : | Yes |
| End of product life : | Recyclable |

PACKAGING INFORMATION

| | |
|--------------------------------------|---|
| Outer box dimension : (W x D x H) | 16.38 po x 9.63 po x 9.63 po 42 cm x 24 cm x 24 cm |
| Box weight: | 18.00 LB / 8.16 KG |
| Conditioning: | 1000 (2 x 500) |
| Storage condition: | Store in a dry place at room temperature |

OTHER

AVAILABLE DOCUMENTS

| | |
|--------------------------|---------------------------|
| Data Sheet | Certificate of Compliance |
| Certificate of Analysis | Safety Data Sheet (SDS) |
| Certificate of Sterility | Pyrogen Declaration |
| DNase/RNase | |

Reach out to us for additional resources, if applicable to this product.

DECLARATION

| | |
|------------|--|
| CFIA | LABPLAS sampling bags are a solution that may be used in the CFIA Preventive Control Plan (PCP) for the seven principles of the HACCP system. The PCP is a Canadian federal initiative, under the Safe Food for Canadians Regulations (SFCR). |
| EU | The materials used to manufacture LABPLAS sampling bags meet, where applicable, the Eu No10/2011 standards for food contact with respect to particle migration. |
| DNase-free | This product is DNase-free. Sensitivity of 10 ⁻⁷ Kunitz units/μL |
| RNase-free | This product is RNase-free. Sensitivity of 10 ⁻⁹ Kunitz units/μL. |
| FDA | The plastic film used in the manufacture of the LABPLAS sampling bag meets the requirements of 21 CFR 177.1520 of the Food and Drug Administration. |
| Pyrogens | This product is non-pyrogenic at the endotoxin limit of 2.15 EU/device. Non-pyrogenicity is supported by endotoxin testing of randomly selected samples using the Limulus amoebocyte lysate (LAL) gel assay according USP-NF <85> and <161> guidelines. |
| Sterile | Sterility is provided by dry heat during extrusion of the plastic at temperatures exceeding 428 F. The approach ensures a sterility assurance level (SAL) of 10 ⁻³ . Continued process effectiveness is demonstrated through periodic sterility testing. Sterility testing follows the USP-NF <71> guideline. |