

GENERAL INFORMATION

Category : Sterile sampling kits

Family : Sani-Sponge

Lifespan: 15 months

TECHNICAL DESCRIPTION

This kit consists of a Twirl'em bag and a biocide-free sponge for sampling various surfaces.



Volume guide

¹Total volume: The total volume corresponds to the maximum capacity the bag can hold when it is filled to the brim, without the possibility of closing it.

²Closure volume: The closure volume corresponds to the maximum capacity the bag can hold when it is closed with a triple twist to ensure safe transport.

³Functional volume: The functional volume corresponds to the volume of diluent or sample according to the test requirements.

SPECIFIC INFORMATION

CONTAINER

Type of container :	Bag
Material :	Thermoplastic laminate
Color :	Transparent
Dimension :	191 x 292 mm / 7.5 x 11.5 in
Thickness :	3.5 mil.in / 0.0889 mm / 88.9 micron
Total volume ¹ :	2100 ml / 70 oz
Closure volume ² :	1200 ml / 41 oz
Functional volume ³ :	100 ml / 3 oz
Printing type :	Writing area
Opening system :	Laser scoring line
Closing system :	Attachment with 2 round wires
Sterile :	Yes
End of product life :	Non-recyclable

INSERTION

Insertion type :	Polyurethane sponge
Description :	Hydrophilic polyurethane foam sponge. White colour. Dimensions (L x W): 3" x 1.25" (77 mm X 31.8 mm). Free from biocides.
Biodegradable :	Non-recyclable

INSERTION

Insertion type :	
Description :	
Biodegradable :	

SOLUTION

Name :	Dey-Engley Neutralizing Broth (DE)
Type of solution :	Transport medium, neutralizing
pH :	7.60±0.2

OTHER

--	--

AVAILABLE DOCUMENTS

Data Sheet	Certificate of Compliance
Certificate of Analysis	Safety Data Sheet (SDS)
Certificate of Sterility	Pyrogen Declaration
DNase/RNase	

PACKAGING INFORMATION

Outer box dimension : (W x D x H)	9.50 X 14.13 X 16.25
Box weight :	6.00 LB / 2.72 KG
Conditioning :	100 (5 x 20)
Storage condition :	Upon receipt, store between 2 °C and 8 °C

Reach out to us for additional resources, if applicable to this product.

DECLARATION

Biocide-free	Biocide-free material
Sterile	Sterility is provided by irradiation. The sterilization dose ensuring a sterility assurance level (SAL) of 10 ⁻³ has been established according to ISO 11137.

1951, rue Nobel, Sainte-Julie, Québec, Canada, J3E 1Z6

Tél : 1.450.649.7343 | Fax : 1.450.649.3113

labplas@labplas.com | www.labplas.com