

GENERAL INFORMATION

Category : Sterile sampling kits

Family : Sani-S'wipe

Lifespan: 5 years

TECHNICAL DESCRIPTION

This kit consists of a sterile bag with an integrated closure system and a biocide-free cloth for sampling various surfaces.



Volume guide

¹Total volume: The total volume corresponds to the maximum capacity the bag can hold when it is filled to the brim, without the possibility of closing it.

²Closure volume: The closure volume corresponds to the maximum capacity the bag can hold when it is closed with a triple twist to ensure safe transport.

³Functional volume: The functional volume corresponds to the volume of diluent or sample according to the test requirements.

SPECIFIC INFORMATION

CONTAINER

Type of container :	Bag
Material :	Thermoplastic laminate
Color :	Transparent
Dimension :	191 x 292 mm / 7.5 x 11.5 in
Thickness :	3.5 mil.in / 0.0889 mm / 88.9 micron
Total volume ¹ :	2100 ml / 70 oz
Closure volume ² :	1200 ml / 41 oz
Functional volume ³ :	250 ml / 8 oz
Printing type :	Form
Opening system :	Laser scoring line
Closing system :	Attachment with 2 round wires
Sterile :	Yes
End of product life :	Non-recyclable

INSERTION

Insertion type :	Blue Cloth
Description :	Non-woven textile composed of cellulose and polyester. Color: pale blue. Folded in 4. Dimensions 9" x 9" (22.9 cm x 22.9 cm). Size (total area): 81 sq. in. (522.5 cm ²). Free from biocides
Biodegradable :	Non-recyclable

INSERTION

Insertion type :	
Description :	
Biodegradable :	

SOLUTION

Name :	NA
Type of solution :	NA
pH :	NA

OTHER

--

AVAILABLE DOCUMENTS

Data Sheet	Certificate of Compliance
Certificate of Analysis	Safety Data Sheet (SDS)
Certificate of Sterility	Pyrogen Declaration
DNase/RNase	

PACKAGING INFORMATION

Outer box dimension : (W x D x H)	9.63 X 16.38 X 9.63
Box weight :	4.00 LB / 1.81 KG
Conditioning :	100 (1 x 100)
Storage condition :	Store in a dry place at room temperature

Reach out to us for additional resources, if applicable to this product.

DECLARATION

FDA	The plastic film used in the manufacture of LABPLAS sampling bags complies with 21 CFR 177.1520 [(c) 3.2c] of the Food and Drug Administration. Its use is limited to temperatures below 212 F according to Table 2 of 21 CFR 176.170 (c).
Sterile	Sterility is provided by irradiation. The sterilization dose ensuring a sterility assurance level (SAL) of 10 ⁻³ has been established according to ISO 11137.