

GENERAL INFORMATION

Category:
Sterile sampling bags

Family: Secure-T

Lifespan: 5 years

TECHNICAL DESCRIPTION

Secure-T bags are made of a flexible, strong and transparent plastic. They have a detachable closure strip that ensures the sterility of the bag.



Volume guide

¹Total volume: The total volume corresponds to the maximum capacity the bag can hold when it is filled to the brim, without the possibility of closing it.

²Closure volume: The closure volume corresponds to the maximum capacity the bag can hold when it is closed with a triple twist to ensure safe transport.

³Homogenizer volume: The homogenizer volume corresponds to the maximum capacity the bag can hold without the risk of overflow when using the homogenizer.

SPECIFIC INFORMATION

ITEM

| ITEM | Bag |
|-----------------------------------|------------------------------------|
| Material : | Polyethylene blend |
| Color : | Transparent |
| Dimension : | 762 x 762 mm / 30 x 30 in |
| Thickness : | 3 mil.in / 0.0762 mm / 76.2 micron |
| Total volume ¹ : | 97000 ml / 3300 oz |
| Closure volume ² : | 53000 ml / 1800 oz |
| Homogenizer volume ³ : | 31000 ml / 1050 oz |
| Printing type : | Clear |
| Opening system : | Perforated line |
| Closing system : | NA |
| Sterile : | Yes |
| End of product life : | Recyclable |

PACKAGING INFORMATION

| | |
|--------------------------------------|---|
| Outer box dimension : (W x D x H) | 17.25 po x 13.63 po x 14.00 po 44 cm x 35 cm x 36 cm |
| Box weight: | 45.20 LB / 20.50 KG |
| Conditioning: | 250 (1 x 250) |
| Storage condition: | Store in a dry place at room temperature |

OTHER

AVAILABLE DOCUMENTS

| | |
|--------------------------|---------------------------|
| Data Sheet | Certificate of Compliance |
| Certificate of Analysis | Safety Data Sheet (SDS) |
| Certificate of Sterility | Pyrogen Declaration |
| DNase/RNase | |

Reach out to us for additional resources, if applicable to this product.

DECLARATION

| | |
|------------|--|
| CFIA | LABPLAS sampling bags are a solution that may be used in the CFIA Preventive Control Plan (PCP) for the seven principles of the HACCP system. The PCP is a Canadian federal initiative, under the Safe Food for Canadians Regulations (SFCR). |
| EU | The materials used to manufacture LABPLAS sampling bags meet, where applicable, the Eu No10/2011 standards for food contact with respect to particle migration. |
| DNase-free | This product is DNase-free. Sensitivity of 10 ⁻⁷ Kunitz units/μL |
| RNase-free | This product is RNase-free. Sensitivity of 10 ⁻⁹ Kunitz units/μL. |
| FDA | The plastic film used in the manufacture of the LABPLAS sampling bag meets the requirements of 21 CFR 177.1520 of the Food and Drug Administration. |
| Pyrogens | This product is non-pyrogenic at the endotoxin limit of 2.15 EU/device. Non-pyrogenicity is supported by endotoxin testing of randomly selected samples using the Limulus amoebocyte lysate (LAL) gel assay according USP-NF <85> and <161> guidelines. |
| Sterile | Sterility is provided by dry heat during extrusion of the plastic at temperatures exceeding 428 F. The approach ensures a sterility assurance level (SAL) of 10 ⁻³ . Continued process effectiveness is demonstrated through periodic sterility testing. Sterility testing follows the USP-NF <71> guideline. |