

## GENERAL INFORMATION

Category:  
Sterile sampling bags

Family: Secure-T

Lifespan: 5 years

## TECHNICAL DESCRIPTION

Secure-T bags are made of a flexible, strong and transparent plastic. They have a detachable closure strip that ensures the sterility of the bag.



## Volume guide

<sup>1</sup>Total volume: The total volume corresponds to the maximum capacity the bag can hold when it is filled to the brim, without the possibility of closing it.

<sup>2</sup>Closure volume: The closure volume corresponds to the maximum capacity the bag can hold when it is closed with a triple twist to ensure safe transport.

<sup>3</sup>Homogenizer volume: The homogenizer volume corresponds to the maximum capacity the bag can hold without the risk of overflow when using the homogenizer.

## SPECIFIC INFORMATION

## ITEM

ITEM	Bag
Material :	Polyethylene blend
Color :	Transparent
Dimension :	140 x 305 mm / 5.5 x 12 in
Thickness :	3 mil.in / 0.0762 mm / 76.2 micron
Total volume <sup>1</sup> :	1400 ml / 47 oz
Closure volume <sup>2</sup> :	840 ml / 28 oz
Homogenizer volume <sup>3</sup> :	590 ml / 20 oz
Printing type :	Clear
Opening system :	Perforated line
Closing system :	NA
Sterile :	Yes
End of product life :	Recyclable

## PACKAGING INFORMATION

Outer box dimension : (W x D x H)	16.38 po x 9.63 po x 9.63 po 42 cm x 24 cm x 24 cm
Box weight:	15.00 LB / 6.80 KG
Conditioning:	1000 (2 x 500)
Storage condition:	Store in a dry place at room temperature

## OTHER

AVAILABLE DOCUMENTS

Data Sheet	Certificate of Compliance
Certificate of Analysis	Safety Data Sheet (SDS)
Certificate of Sterility	Pyrogen Declaration
DNase/RNase	

Reach out to us for additional resources, if applicable to this product.

DECLARATION

CFIA	LABPLAS sampling bags are a solution that may be used in the CFIA Preventive Control Plan (PCP) for the seven principles of the HACCP system. The PCP is a Canadian federal initiative, under the Safe Food for Canadians Regulations (SFCR).
EU	The materials used to manufacture LABPLAS sampling bags meet, where applicable, the Eu No10/2011 standards for food contact with respect to particle migration.
FDA	The plastic film used in the manufacture of the LABPLAS sampling bag meets the requirements of 21 CFR 177.1520 of the Food and Drug Administration.
Sterile	Sterility is provided by irradiation. The sterilization dose ensuring a sterility assurance level (SAL) of 10 <sup>-3</sup> has been established according to ISO 11137.